

CALIBRATION RECORD OF HARDNESS TESTER

Page No.: 1 of 1

Make		Location	IPQA Lab.
ID No.		Frequency	Monthly
Calibration done on		Next Due Date	

Weight Box ID: _____ **Calibration Done On:** _____ **Due On:** _____

OBSERVATIONS

Entered Weight (kg)	Measured Weight (kg)	Acceptance Limit (Tolerance Limit: $\pm 5\%$)	OK / Not OK	Done By (Sign/Date)
01 kg		0.95 – 1.05 kg		
02 kg		1.90 - 2.10 kg		
05 kg		4.75 – 5.25 kg		
10 kg		9.50 – 10.50 kg		
20 Kg		19.00 – 21.00 kg		

Note: Use 01 kg, 02 kg, 5 kg, 10 kg and 20 kg calibrated weights to validate the load Cell.

Remarks: The Hardness Tester is calibrated and is qualified / not qualified for use.

Verified by: _____
(Sign/Date)