

CALIBRATION RECORD OF LEAK TEST APPARATUS

Page No.: 1 of 1

Name	Leak Test Apparatus	Location	
Make / Model		Frequency	Monthly
Serial No.		Calibration done on	
ID No.		Next Due Date	

Sr. No.	Set Time (In seconds)	Observed Time (In seconds)	Acceptance Limit \pm 10 Seconds	Done By (Sign.)
1.	60 seconds		50 seconds – 70 seconds	
2.	120 seconds		110 seconds – 130 seconds	
3.	180 seconds		170 seconds – 190 seconds	
4.	300 seconds		290 seconds – 310 seconds	

Remarks: The Leak Test Apparatus is calibrated and is Qualified / Not Qualified for use.

Done By	Checked By	Verified By (QA)
(Sign./Date)	(Sign./Date)	(Sign./Date)